

Confirmation 21449391a

Airborne room disinfection according to EN 17272*

* EN 17272- Chemical disinfectants and antiseptics —
Methods of airborne room disinfection by automated process —
Determination of bactericidal, mycobactericidal, sporicidal, fungicidal, yeasticidal, virucidal and phagocidal activities

The above mentioned and for this purpose accredited testing laboratory confirms that the performed disinfection procedure

gassing with the ozone generator from the company Oxy3

by the company

Oxy3 Ozongeräte Produktion GmbH

Parkring 1

8712 Niklasdorf

Österreich

has shown

**a sufficient bactericidal effect
under clean conditions (low organic load)
for the medical area
in a small chamber (<2m³) and
an exposure time of 90 minutes**

in accordance with the requirements according to EN 17272.

Details can be found in test report 21449301 and expert opinion 21449381.

This confirmation is not proof of complete testing and complete fulfillment of all requirements from EN 17272.

Bischofshofen, 06.04.2022


Dr. Arno Sorger
Technical manager W.H.U. GmbH